

# **Hercules**<sup>TM</sup> **Balloon**

Balloon Inflation Catheter

# Product Overview

## Model Description


Reference	Maximum inflation diameter (mm)	Catheter working length (mm)
HD40-100	40mm	100cm

## Specifications

Balloon Diameter (D) and corresponding liquid volume (V)	D	V
	26mm	12 ± 2ml
	31mm	18 ± 2ml
	36mm	24 ± 3ml
	40mm	34 ± 3ml
Guidewire compatibility	0.035"	
Shaft profile	10F	
Recommended Introducer sheath	14F	

## Materials

Balloon	Polyurethane (PU)
Tube	Polyurethane (PU)
Marker Band	Platinum-iridium (Pt/Ir)
Stopcock	Polycarbonate (PC)
Y-Connector	Polycarbonate (PC)



Lombard Medical Limited, Lombard Medical House, 4 Trident Park, Didcot, OX11 7HJ  
 Tel: +44 (0)1235 750800 Email: [globalinfo@endovastec-lm.com](mailto:globalinfo@endovastec-lm.com) Website: [www.lombardmedical.com](http://www.lombardmedical.com)

All cited trademarks are the property of their respective owners in respective countries. These devices should only be used by a licenced physician. Indicators, contraindications, warnings, potential adverse events and directions for use can be found in Instructions for Use and product labelling supplied with each device. The promotion, distribution or sale of these devices is prohibited in countries where the registration certificates or any other licences have not been obtained. This brochure is for the use of the individuals to whom it has been provided and is not to be distributed widely as an advertisement. Any information may be changed without prior notice. Hercules Balloon™ is manufactured by MicroPort Endovastec and Distributed in Europe by Lombard Medical Limited. Not Available for sale in the United States. ©2026 Lombard Medical Limited. All rights reserved. D01519\_ENG A